

Digital Radiography System

DR-300

IHE Integration Statement

REVISION HISTORY

REVISION	DATE	COMMENTS
First Edition	2015/04	Document is created.

Contents

1. Introduction	1
2. Connectathon	1
3. IHE Integration Statement.....	2

This page is intentionally left blank.

1. Introduction

This manual describes the Digital Radiography unit DR-300 conformity to IHE (Integrating the Healthcare Enterprise).

IHE shows the ideal healthcare environment in the hospital.

To achieve the implementation along with IHE, Shimadzu have a willing to provide an ideal operation environment for our customers.

2. Connectathon

There is the test event called "Connectathon", is held in order to verify systems compliant with the IHE. Shimadzu had attended the North America Connectathon in 2014, and passed the prescribed test.

Note1) DR-300 is not tested in other regional IHE Connectathon such as IHE-J than IHE-NA.

Note2) DR-300 conforms to IHE from Rev.01.05.00.

Please refer to the following site, to see the Connectathon test results.

<http://connectathon-results.ihe.net/>

Please refer to the following site, to see the IHE information of Shimadzu.

<http://www.shimadzu.com/products/medical/index.html>

IHE Integration Statement			
Vendor	Product Name	Version	Date
Shimadzu Corp	MOD_SHIMADZU_RF	Rev.01.05.00	4/30/15

This product implements all transactions required in the IHE Technical Framework to support the IHE Integration Profiles, Actors and Options listed below :

Integration Profile	Actor	Integration Profile Option
Radiation Exposure Monitoring	Acquisition Modality	None
Scheduled Workflow	Acquisition Modality	Patient Based Worklist Query
Scheduled Workflow	Acquisition Modality	None

Internet address for vendors IHE information : http://www.shimadzu.com/med/products/dicomihe/index.html	
Links to Standards Conformance Statements for the Implementation	
HL7	
DICOM	http://www.shimadzu.com/med/products/fluoro/index.html
Link to general information on IHE : www.ihe.net	